

⚠ WARNING: Connection to the gas source must be done by using only the appropriately indexed fitting.

4. Ensure that the Float Ball is at the very bottom of the Flow Tube when turned off.

NOTE: If the Float Ball is not at the bottom of the Flow Tube, the Flowmeter could be leaking.
Please contact your dealer or Ohio Medical Corporation.

5. Adjusting the Flow: To DECREASE Flow: Turn the knob clockwise
To INCREASE Flow: Turn the knob counter clockwise

6. To set the flow, align the center of the float ball to the indicator line on the Flow Tube.

⚠ WARNING: To avoid injury ALWAYS confirm flow requirement for patient prior to dispensing.
Check flow frequently while being administered to patient.

7. If the flow is adjusted beyond the highest calibrated indicator, an undetermined flow will arise.
8. Turn knob completely counter clockwise to achieve maximum flood/ flush flow.

NOTE: Any flow beyond the highest calibrated line on the Flow Tube with unrestricted flow is Flood/Flush flow.

Cleaning Instructions

Use a clean damp cloth with a mild cleaning solution to wipe outside of product. DO NOT gas sterilize with Ethylene Oxide (ETO). DO NOT clean with pungent hydrocarbons.

CAUTION: DO NOT submerge product in any form of liquid. This will cause damage and void the warranty.

Troubleshooting

If the Flowmeter does not function properly, contact your dealer or equipment provider.

Maintenance Prevention

Before and after each use, INSPECT THE FLOWMETER FOR DAMAGE.

⚠ WARNING: When changing connectors on the Flowmeter for service or replacement, NEVER re-attach connectors of a different gas. Doing so may result in patient injury or damage to the equipment.

CAUTION: Disconnect Flowmeter from gas supply BEFORE SERVICING.

⚠ WARNING: Ensure shroud is securely tightened to prevent disconnection under pressure.
Turn Clockwise to tighten.

WARRANTY

7700 Series Flowmeter here known as "Product" is sold by Ohio Medical Corporation, a Delaware corporation (the "Company") under the express terms of the warranty set forth below.

For a period of SIXTY (60) MONTHS [with the exception of the 0-200 CCM and 0-1 LPM Flowmeter for TWELVE (12) MONTHS] from the date the Company ships, this Product is warranted to be free from functional defects in materials and workmanship and to conform in all material respects to the description for the Product contained in this operation manual, if this Product is properly operated under conditions of normal use, regular periodic maintenance and service is performed and repairs are made in accordance with this operation manual.

The Company's sole and exclusive obligation and customer's sole and exclusive remedy under the above warranty is limited to repair or replacement, at the Company's option, of the defective Product.

The foregoing warranty shall not apply if the Product has been repaired or altered by anyone other than the Company or an authorized dealer; or if the Product has been subjected to abuse, misuse, negligence, or accident.

The Company reserves the right to stop manufacturing any product or change materials, designs, or specifications without notice.

This warranty is extended to only the initial customer with respect to the purchase of this Product directly from the Company or an authorized dealer as new merchandise. Dealers are not authorized to alter or amend the warranty of any Product described in this agreement. Any statements, whether written or oral, will not be honored or be made part of the agreement of sale.

THIS WARRANTY IS EXPRESSLY IN LIEU OF ANY OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE. THE COMPANY SHALL NOT BE LIABLE FOR INCIDENTAL, COLLATERAL, CONSEQUENTIAL, OR SPECIAL DAMAGES INCLUDING, BUT NOT LIMITED TO, LOST PROFITS, OR LOSS OF USE. THE COMPANY'S LIABILITY, IN THE AGGREGATE, SHALL NOT EXCEED THE PURCHASE PRICE OF THE PRODUCT.

In order to file a warranty claim, customer is required to return Product prepaid to the Company at 1111 Lakeside Drive, Gurnee IL, 60031 USA. As determined at the sole discretion of the Company, Products which qualify under the warranty will be repaired or replaced, at the Company's option, and returned to customer via ground delivery at the Company's expense.

All claims for warranty must first be approved by Ohio Medical Corporation Customer Service Department. For US Domestic customer returns: customer.service@ohiomedical.com or 1-800-662-5822 (Option 3). For Canadian customer returns: intl.customerservice@ohiomedical.com or 1-800-662-5822 (Option 3). Upon approval the customer service department will issue a Return Goods Authorization (RGA) number. An RGA must be obtained prior to commencement of any warranty claim.

OPERATING AND MAINTENANCE MANUAL

7700 OXYGEN / MEDICAL AIR FLOWMETER



Flowmeters shown with optional Clear Nut & Nipple

Rx ONLY

© 2013 Ohio Medical Corporation. This document contains information that is proprietary and confidential to Ohio Medical Corporation. Use of this information is under license from Ohio Medical Corporation. Any use other than that authorized by Ohio Medical Corporation is prohibited. Ohio Medical Corporation and the Ohio Medical logo are registered trademarks of Ohio Medical Corporation.

Ohio
Medical Corporation®
always there | for life

1111 Lakeside Drive, Gurnee, IL 60031-4099
Tel: 1-866-549-6446 | Fax: 847-855-6218
www.ohiomedical.com

IMPORTANT: SAFETY INSTRUCTIONS

This manual provides you with important information about the Flowmeter and should be read carefully to ensure the safe and proper use of this product.


READ AND UNDERSTAND ALL THE SAFETY AND OPERATING INSTRUCTIONS CONTAINED IN THIS BOOKLET BEFORE USING THIS PRODUCT.

IF YOU DO NOT UNDERSTAND THESE INSTRUCTIONS, OR HAVE ANY QUESTIONS, CONTACT YOUR SUPERVISOR, DEALER OR THE MANUFACTURER BEFORE ATTEMPTING TO USE THE APPARATUS.

⚠ WARNING: Indicates a potentially hazardous situation, if not avoided, could result in death or serious injury.

ATTENTION: Indicates a potentially hazardous situation, if not avoided, could result in minor or moderate injury.

CAUTION: Indicates a potentially hazardous situation, which, if not avoided could result in damage to the device or other property.



Caution, consult accompanying documents

Use no oil

Read Lpm
SN Serial Number

Receiving Inspection

Remove product from package and inspect for damage. If product is damaged, DO NOT USE and contact your dealer or equipment provider.

CAUTION: The tube inside the 0-200 CCM and 0-1 LPM flowmeter is made of glass and is fragile. Special care should be taken to avoid breaking the glass.

ATTENTION: It is very important to allow product to remain in original packaging for 24 hours to acclimate to room temperature before use.

ATTENTION: Store the product in a sealed package to avoid environmental damage. The operating and storage temperature for the Flowmeter should reflect typical environmental conditions of a medical facility environment.

User Responsibility

⚠ WARNING: Service of this device should ONLY be performed by properly trained individuals. The Flowmeter must ONLY be used under the direction of a healthcare professional.

This product performs as explained in this manual, as long as the assembly, use, repair and maintenance are properly followed according to our instructions. Periodic review of this device is recommended. If any damage or defects are present, the product should not be used. This includes parts that may have been altered, contaminated, worn or missing. If any of the above are noted, immediate repair / replacement is required. In compliance with the manufacturers warranty, repair of this device is not to be performed by anyone other than a qualified professional. If this device is subject to improper maintenance, repair, use and/or abuse leading to malfunction of the device, replacement is the sole responsibility of the user. In compliance with the Ohio Medical warranty, repair of this device is not to be performed by anyone other than an Ohio Medical trained professional and done in strict accordance to the written instructions provided by Ohio Medical. If this device is subject to improper maintenance, repair, use and or abuse leading to malfunction of the device, replacement is the sole responsibility of the user.

⚠ WARNING: Due to the possibility of explosion caused by static charge, operation of this device is not to be done if flammable anesthetics are present.

TO MINIMIZE THE RISK OF EXPLOSION OR FIRE:

- NEVER attempt to attach a Flowmeter directly to a cylinder.
- NEVER use grease, oil, organic lubricants or flammable materials on or near the Flowmeter.
- NEVER smoke in an area where oxygen is being used.
- NEVER use any type of flame or flammable or explosive material near the Flowmeter.

- ALWAYS follow CGA and ANSI standards for Flowmeters and Medical Gas Products (E-7) and Oxygen Handling (G-4).

ATTENTION: Keep the Flowmeter in a clean area when not being used.

ATTENTION: Ensure that all connections are tightened and free of leaks prior to use. Use only an oxygen-safe leak detector when testing for leaks.

⚠ WARNING: Each Flowmeter can only be used with one type of gas.

Intended Use

A compensated thorpe tube flowmeter is a device intended for medical purposes that is used to control and measure gas flow rate accurately. The device includes a vertically mounted tube, with the outlet of the flowmeter calibrated to a reference pressure.

Specifications

Gas	Scale	Increments	Accuracy Averages	Max. Flood/Flush	Transport/Storage Requirements
Oxygen	0-200 CCM	25 CCM (starts at 25 CCM)	+/- 20 CCM	500 CCM	-40°F to 140°F (-40°C to 60°C)
Oxygen	0-1 LPM	0.1 LPM (starts at 0.1 LPM)	+/- 0.1 LPM	2.5 LPM	-40°F to 140°F (-40°C to 60°C)
Oxygen	0-3.5 LPM	0.125 LPM (from 0.125 to 1 LPM) 0.25 LPM (from 1 to 3.5 LPM)	+/- 0.125 LPM or +/- 10% of reading (whichever is greater)	45 LPM	-40°F to 140°F (-40°C to 60°C)
Gas	Scale	Increments	Accuracy Averages	Min. Flood Flush	Transport/Storage Requirements
Oxygen & Medical Air	0-15 LPM	0.5 LPM (from 1 to 5 LPM) 1 LPM (from 5 to 15 LPM)	+/- 0.5 LPM or +/- 10% of reading (whichever is greater)	50 LPM	-40°F to 140°F (-40°C to 60°C)

⚠ WARNING: Prior to administering medical air or oxygen, confirm required flow and frequently monitor the flow.

⚠ WARNING: 0-200 CCM, 0-1 LPM and 0-3.5 LPM Flowmeters are NOT to be used for resuscitation.

Flowmeters are calibrated at the pressure indicated on the flow tube at 70° F (21° C), standard atmospheric pressure. Specifications are subject to change without prior notice.

⚠ MRI WARNING: This product contains magnetic, ferrous material that may affect the result of an MRI. MR conditional options may be available.

Operating Instructions

1. Turn Flowmeter off by turning knob fully clockwise.
2. Inspect the Flowmeter for damage. If any damage is found, DO NOT use the Flowmeter.

CAUTION: Over tightening the knob when turning the Flowmeter off will cause damage. The Flowmeter must be used with the Flow Tube in an upright position.

⚠ WARNING: The Flow Tube specifies the gas and pressure required. The accuracy will be affected if a different pressure, other than the one specified, is used.

3. Connect the Flowmeter to the supply pressure specified on the Flow Tube.

⚠ WARNING: The accuracy may be affected if the temperature of the gas is different than 70° F (21° C) and the supply pressure is different than that indicated on the Flow Tube.

NOTE: The accuracy of the flow will not be affected by the attachment of accessories, however, the indicated flow may change.