

Vacuum Tester

DESCRIPTION

Vacuum system flow rate and negative pressure test device with stand and carrying case; Analog vacuum gauge, 0-760 mmHg and 0-30 in Hg negative pressure; flow scale 15-150 L/min air.

Part Number: 0175-1453-000



SPECIFICATIONS

Flowmeter (High Flow version)

High Flow: 15-150 L/min (0.5 to 5.3 SCFM)
 Accuracy: +/- 5% of full scale
 Float: Carbon carbide with flat edge plus guide rod (take reading at the top edge of the float)
 Two quick-reference arrows, for marking flow rate ranges

Vacuum Gauge

Range: 0-760 mmHg (0-30 inHg)
 Accuracy: +/- 5% OF FULL SCALE

Service Parts:

Gauge: AI5110011
 Base: AI5110021R

Vacutimer Portable Digital Timer

DESCRIPTION

The Vacutimer is a portable, dual-display digital timer designed expressly for setting or checking the timing of the Intermittent Suction Units. The Vacutimer frees the USER to make adjustments to one half cycle while the other is being timed. When utilized in this fashion, the Vacutimer will yield the user a time savings of greater than 50% over any other method.

Part Number: 6700-0438-800



FEATURES:

- Automatic operation
- Easy to use
- Employs no proprietary components
- Unique dual digital display
- Crystal-referenced time base
- Complete with carrying case and accessories
- 9V Battery Powered (not included)

Ohio Medical Corporation and the Ohio Medical Corporation logo are registered trademarks of Ohio Medical Corporation.

TRUSTED BRANDS OF OHIO MEDICAL®



Ohio Medical Corporation®
 1111 Lakeside Drive - Gurnee, IL 60031 USA
 TollFree: 866-549-6446 - Fax: 847-855-6218
 www.ohiomedical.com
 255494 Rev.2 (04/2014)